



Tentative Interim Amendment

## NFPA<sup>®</sup> 99

### *Health Care Facilities Code*

#### 2018 Edition

**Reference:** 14.2.10.1.2

**TIA 18-1**

(SC 17-8-56 / TIA Log #1241)

**Note:** Text of the TIA was issued and approved for incorporation into the document prior to printing.

*1. Revise paragraph 14.2.10.1.2 to read as follows:*

**14.2.10.1.2** Detectors, sensors, transducers, and communications equipment located inside the chamber shall meet the requirements of paragraph 14.2.9.3.16 for Class A chambers and paragraph 14.2.9.6 for Class B chambers.

**Issue Date:** August 17, 2017

**Effective Date:** September 6, 2017

(Note: For further information on NFPA Codes and Standards, please see [www.nfpa.org/docinfo](http://www.nfpa.org/docinfo))

Copyright © 2017 All Rights Reserved  
NATIONAL FIRE PROTECTION ASSOCIATION